

**METER ERROR REPORT****FINAL**

Reconcile?	Y
------------	---

Safety Issue?	Y/N
---------------	-----

**1. EXECUTIVE SUMMARY**

SITE NAME	Dowlais	
LDZ	WS (Wales & West)	
START DATE (actual)	12 <sup>th</sup> August 2015	
LAST GOOD DATE		
END DATE	13 <sup>th</sup> November 2015	
SIZE OF ERROR (No reconciliation required if under 0.1%)	0.4897% over registration (10.79 GWh)	
ESTIMATE – Y/N?		
ROOT CAUSE	Incorrect K Factors used in flow computer for stream 2	
ANALYSIS	Recalculation of volumes using correct K Factors	
METER TYPE	USM	
AUTHOR	C. Litster	
CHECKED BY		
ACCEPTED BY NETWORK		
RECONCILIATION	Distribution	Transportation

## 2. BACKGROUND

Dowlais has a duty/standby ultrasonic meter streams using a gas chromatograph for CV determination and PTZ correction.

During the period from 11:13 on the 12<sup>th</sup> of August 2015 to 13:45 on the 13<sup>th</sup> November 2015 the incorrect K factors were used in the flow computer. Meter 07278615-1115 in metering stream 2 was calibrated at DNV-GL Flow Centre on the 10<sup>th</sup> August 2015, when the meter was returned to use on the 12<sup>th</sup> August 2015 the K factors were not updated in the flow computer to reflect the new certificate.

## 3. ERROR QUANTIFICATION AND IMPACT

The archived metering and gas quality data was downloaded from HPMIS. Using RBD data and the incorrect K factors the incorrect actual m<sup>3</sup> and the standard m<sup>3</sup> flow rates were calculated. This was repeated using the correct K factors to recalculate the corrected actual m<sup>3</sup> and the standard m<sup>3</sup> flow rates. The error was calculated on a daily basis as the difference between actual m<sup>3</sup> and the standard m<sup>3</sup> flow rates using the corrected input.

The overall error is an over-registration of 0.4897%. The error is equivalent to 10.79 GWh.

Freqtest (Hz)	Qref (m3/h)	Error (%)	Kfactor (pls/m3)
1730.20	12140.00	-0.18	513.07
1212.18	8508.65	-0.22	512.87
670.83	4708.38	-0.21	512.92
425.78	2988.79	-.022	512.86
174.76	1224.95	-0.08	513.61
88.17	618.14	-0.10	513.48

**Table 1 – Dowlais Ultrasonic Meter Factors 2007, Certificate No. 04037**

Freqtest (Hz)	Qref (m3/h)	Error (%)	Kfactor (pls/m3)
1719.51	12002.14	0.34	515.76
1286.26	8976.52	0.36	515.85
816.45	5696.64	0.38	515.95
443.33	3094.13	0.35	515.81
168.89	1178.36	.039	515.99
90.63	631.62	0.50	516.57

**Table 2 – Dowlais Ultrasonic Meter Factors 2015, Certificate No. 11666-1**

#### 4. CAUSES

Meter 07278615-1115 in metering stream 2 was calibrated at DNV-GL Flow Centre on the 10<sup>th</sup> August 2015, when the meter was returned to use on the 12<sup>th</sup> August 2015 the K factors were not updated in the flow computer to reflect the new certificate.

#### 5. RECOMMENDATIONS AND LEARNING

The error should be reconciled by applying the offtake daily flows detailed in Appendix A.

On the 31<sup>st</sup> October 2015 and 1<sup>st</sup> – 3<sup>rd</sup> November 2015 the Relative Density and Calorific Value were missing due a fault with the chromatograph, therefore a correction has been applied equal to the mean correction factor for the previous and subsequent day.

#### REFERENCES

HPMIS Database, DNV- GL Meter Calibration Certificate 04037, DNV- GL Meter Calibration Certificate 11666-1.

#### VERSION HISTORY

<i>Version</i>	<i>Changes</i>	<i>Author</i>	<i>Date</i>
0	Original	C. Litster	25/11/2015

#### DISTRIBUTION

*Wales & West Utilities*

#### APPENDIX A – Daily Corrected Flows

The table below gives the daily correction factors which should be used to reconcile the error.

<b>Gas Day</b>	<b>Daily Correction Factor</b>
12/08/2015	0.994199
13/08/2015	0.994332
14/08/2015	0.995071
15/08/2015	0.994575
16/08/2015	0.994971
17/08/2015	0.994402
18/08/2015	0.994649
19/08/2015	0.994799
20/08/2015	0.994957
21/08/2015	0.994488
22/08/2015	0.994724

<b>Gas Day</b>	<b>Daily Correction Factor</b>
28/09/2015	0.995123
29/09/2015	0.995268
30/09/2015	0.995296
01/10/2015	0.995118
02/10/2015	0.995179
03/10/2015	0.994981
04/10/2015	0.994986
05/10/2015	0.995167
06/10/2015	0.995088
07/10/2015	0.995184
08/10/2015	0.995272

23/08/2015	0.995092
24/08/2015	0.995187
25/08/2015	0.995102
26/08/2015	0.994901
27/08/2015	0.994807
28/08/2015	0.994880
29/08/2015	0.994892
30/08/2015	0.994696
31/08/2015	0.995082
01/09/2015	0.995019
02/09/2015	0.994865
03/09/2015	0.995070
04/09/2015	0.995062
05/09/2015	0.994391
06/09/2015	0.994476
07/09/2015	0.994892
08/09/2015	0.994837
09/09/2015	0.994937
10/09/2015	0.994611
11/09/2015	0.994819
12/09/2015	0.994700
13/09/2015	0.995003
14/09/2015	0.995036
15/09/2015	0.995078
16/09/2015	0.995146
17/09/2015	0.995325
18/09/2015	0.995099
19/09/2015	0.994897
20/09/2015	0.995243
21/09/2015	0.995178
22/09/2015	0.995140
23/09/2015	0.995029
24/09/2015	0.995111
25/09/2015	0.995259
26/09/2015	0.995240
27/09/2015	0.995171

09/10/2015	0.994919
10/10/2015	0.995251
11/10/2015	0.995232
12/10/2015	0.995208
13/10/2015	0.995053
14/10/2015	0.995091
15/10/2015	0.994950
16/10/2015	0.995083
17/10/2015	0.994949
18/10/2015	0.995043
19/10/2015	0.995021
20/10/2015	0.995287
21/10/2015	0.995161
22/10/2015	0.995248
23/10/2015	0.994969
24/10/2015	0.994950
25/10/2015	0.995038
26/10/2015	0.995042
27/10/2015	0.995259
28/10/2015	0.995259
29/10/2015	0.995128
30/10/2015	0.995217
31/10/2015	<b>0.995161</b>
01/11/2015	<b>0.995161</b>
02/11/2015	<b>0.995161</b>
03/11/2015	<b>0.995161</b>
04/11/2015	0.995105
05/11/2015	0.995123
06/11/2015	0.995126
07/11/2015	0.995270
08/11/2015	0.995098
09/11/2015	0.995105
10/11/2015	0.995170
11/11/2015	0.995083
12/11/2015	0.994944
13/11/2015	0.994938